

## CORRIGENDUM

Based on pre-bid meeting held on 24-03-2026, the following amendments in GeM Bid No.: **GEM/2026/B/735223 dated: 16-03-2026**, are notified:

| <b>Sr. no.</b> | <b>Existing specifications in the tender</b>  | <b>Revised Specifications</b>  |
|----------------|---|--|
| 1.             | <b>Total Vessel Volume:</b> 10-13 Liters with at least 3 litres of free space above slurry for containing gases produced in Bio-reactor                           | <b>Total Vessel Volume:</b> 13 Liters with at least 3 litres of free space above slurry for containing gases produced in Bio-reactor   |
| 2.             | <b>Working Volume:</b> 7-10 Liters for feed slurry  | <b>Working Volume:</b> 10 Liters for feed slurry   |
| 3.             | <b>Warranty:</b> On-site 3 years comprehensive warranty and 2 years non-comprehensive on complete system, software and electronic accessories should be provided. | <b>Warranty:</b> On-site 1 year comprehensive warranty and 2 years non-comprehensive on complete system, software and electronic accessories should be provided.<br><br>Duration of e-PBG was kept for 14 Months.<br><br>AMC Service for next 2 years for the offered product at the rate of not more than 5% of contract price per annum, which will be paid Bi-annually.<br><br>The original Performance Security of contract will be returned only after submission and verification of AMC Performance Security for 5% of total AMC value valid up to AMC period plus 2 months (if there is no other claim).<br><br>Bid value should include cost of Equipment along with AMC charges with price break-up. |

All other specifications with terms and conditions will remain unchanged as per the original tender.